



Resvent iBreeze Auto CPAP & Bilevel Systems

Introducing the Resvent iBreeze Auto CPAP and Bilevel Systems with Heated Humidification, feature-rich PAP devices that were granted EUA (Emergency Use Authorization) by the FDA in 2020.* Some of the features include IPR (Intelligence Pressure Release), an algorithm during inspiration and expiration offering patients a more comfortable breathing experience; an auto humidifier which adjusts the heat based on room temperature and humidity to prevent rainout in the hose or mask; and finally, a 2-in-1 humidifier design which monitors the water level and when low, will signal a visible alert and automatically switch off the heat to prevent damage to the device. The iBreeze Systems come complete with SD Card, Carry Bag, 6' CPAP Tube, Tube Connector, Power Cord, and User Manual. *Wi-Fi equipped devices optional.*

SPECIFICATIONS

H2O Therapy Pressure:	4-20 cm (APAP) 4-30 cm (BPAP)
Ramp Time:	0-60 minutes (5-minute increments)
Dimensions:	9.25" w x 5.8" d x 4.9" h (device); 14.25" x 9.75" x 7" (shipping)
Weight:	3.4 lbs (device); 6.2 lbs (shipping)
Water Capacity:	290 mL
Humidifier Settings:	Auto / 0-8 (Default: 3)
Humidifier Output:	≥10 mg H2O/L
Warranty:	Two Years Limited (device)



FEATURES & BENEFITS

Large 3.5" LCD Display

Easily adjust settings and view last session to encourage user compliance.

Quiet Operation

with real time audible alert for tubing/mask leaks and automatic leakage/altitude compensation.

Intelligence Pressure Release

with 3 comfort settings and inspiration trigger for automatic startup.

Compliance Data

access compliance reporting via SD card or optional Wi-Fi.

Embedded Memory

stores previous night and 365 night user raw data.

iMatrix/ResAssist

Register your patients to maintain records and generate sleep reports on iMatrix desktop software or ResAssist cloud platform.

Item #	Description	HCPSCS	UOM	Case QTY
IBREEZE20A	iBreeze Auto CPAP System with Heated Humidification	E0601	Each	1
IBREEZE20A-W	iBreeze Auto CPAP System with WiFi and Heated Humidification	E0601	Each	1
IBREEZE30STA	BiLevel Auto S/T System with WiFi and Heated Humidification	E0471	Each	1
Replacement Parts				
05205-000001-00	iBreeze Replacement 2-Piece Power Adapter with Brick	N/A	Each	1
1001-000222-00	iBreeze Replacement Humidifier Chamber	N/A	Each	1
CPF-IBREEZE	iBreeze Replacement Disposable Filter	N/A	Pack	2
69999-000032-00	iBreeze APAP Replacement Knob	N/A	Each	1
69999-000033-00	iBreeze BPAP Replacement Knob	N/A	Each	1
10001-010043-00	iBreeze Replacement Tube Adapter	N/A	Each	1

FAQ's and Troubleshooting

1

How do I access and make changes in the Patient Menu?

Turn the knob until the setting category you wish to adjust is highlighted. Press the knob button to enter. Continue turning the knob and pressing the knob button to make your adjustments.

2

Why can't I access the Humidity menu?

The humidity menu will not enable until there is water in the humidification chamber.

3

It feels like the pressure keeps going up and down. Is there something wrong?

Try turning the IPR setting down or off. Also, make sure the mask has a good seal on the face. If there is any leak, the device will recognize that it needs to accommodate for the lost pressure and increase pressure up.

4

How do I access the Clinician Menu?

Press the Ramp button and knob button simultaneously for at least 3 seconds.

5

How do I upload patient data and access Compliance Reporting?

Follow this link to install iMatrix desktop software: <https://bit.ly/3dPF8I2>

Follow this link to access ResAssist cloud platform: <https://usa.resvent.com/ResAssist/resvent>

6

The SD card is not being recognized in iMatrix. What should I do?

Close iMatrix and remove the SD card from your computer. Re-open iMatrix and reinsert the SD card. Also, try rebooting your computer if this does not work.

7

How do I clean the iBreeze device? Humidifier?

Unplug the device before cleaning. Do not let any liquid enter inside of the device. Clean the outer housing with a clean cloth lightly sprayed or moistened with a mild detergent or disinfecting cleaner. Dry thoroughly before reconnecting to the power supply. To clean the humidifier, open the lid. Fill with warm water with a small amount of mild dish detergent. Wash surfaces with a clean cloth or sponge. Rinse with warm water until all suds are gone and allow to air dry thoroughly.

8

How often should I replace the Filter? Humidifier Chamber? Tubing?

The filter should be changed every 2-4 weeks.

The humidifier chamber and tubing should be replaced every 3-6 months.

9

What does the 2 Year Limited warranty cover?

The warranty covers the internal function of the device and display. The warranty does not include damage caused by accident, abuse, alteration, water ingress, and other any other defects related to the material or workmanship. Accessories such as the device housing, humidification chamber, power cords, filter, tubing, or tubing adapter do not fall under the warranty.

*For more information on the FDA Emergency Use Authorization, click here: [Ventilators and Ventilator Accessories EUAs](#)